

Important information and recommendations regarding the European Medical Device Regulation (MDR)

Once the transition period expires, the new European Medical Device Regulation (MDR) will be implemented from May 26, 2021.

In order to implement the challenges and new requirements as safely as possible together with you, we have put together some information for you below.

What does the new Medical Device Regulation regulate?

The new European Medical Device Regulation defines the requirements regarding conformity assessment of medical devices.

It combines the Medical Devices Directive (93/42/EEC) and the Active Implantable Medical Devices Directive (90/385/EEC) with the objective of implementing a higher standard and further increasing the focus on patient safety.

What is the reason behind the UDI (Unique Device Identification) system?

Under the Unique Device Identification (UDI) system, medical devices must bear a globally unique product number that is machine readable and included on the product and/or packaging, for example in a barcode. This code acts as a key for a UDI database (UDID) that contains various details about the products.

Using the UDI improves the documentation and traceability of the materials used. The number enables information to be traced, such as the manufacturer of the medical device, product, manufacturing date and the production batch.

Documentation on delivery notes or invoices is not required, as the products are frequently removed from storage and the batch number only needs to be documented when processing is carried out.

Implementing the MDR in accordance with VITA quality standards

1. VITA veneering and CAD/CAM materials

All VITA veneering and CAD/CAM materials bear a UDI-compliant Health International Bar Code (HIBC) that contains all the required information, either in digital form or in plain text. All units produced (e.g., bottles or blocks) are labeled with a batch number. The scannable barcode is provided on the primary packaging of the product.

2. VITA prosthetic teeth

Our tooth sets are also labeled with a UDI-compliant Health International Barcode. Every tooth set includes a verified and approved tooth batch that ensures correct batch traceability. As the risk class is low, batch identification is not required for each individual tooth.



Tooth set description

VITAPAN EXCELL, VITAPAN LINGOFORM, VITAPAN PLUS, VITAPAN, VITAPAN CUSPIFORM, VITA PHYSIODENS



Nade in Germany

- 1. Tooth shade
- 2. Tooth line
- 3. Tooth shape/size
- 4. LOT number
- 5. HIBC code

Identified HIBC code:



- 6. VITA manufacturer number
- 7. Product number
- 8. VITA packaging index
- 9. Manufacturing date
- 10. Manufacturer
- 11. Only to be used by technical staff
- 12. CE mark
- 13. Country of manufacture

Tooth set description

VITAPAN SYNOFORM, VITA LUMIN VACUUM, VITA PHYSIODENS



- 1. Tooth shade
- 2. Tooth line
- 3. Tooth shape/size
- 4. Manufacturer
- 5. HIBC code

Identified HIBC code:





- 6. VITA manufacturer number
- 7. Product number
- 8. VITA packaging index
- 9. LOT number
- 10. Control character for manufacturing date
- 11. Manufacturing date
- 12. Check number
- 13. Manufacturer address
- 14. CE mark



VITA tips and recommendations

Documentation options

To facilitate documentation, we recommend using a documentation software. Using a barcode reader, all UDI-compliant HIBC information on the medical devices can be scanned, transferred to the documentation software and saved.

Manual documentation of the product number and batch to ensure sufficient traceability, can also be carried out without a barcode reader.

We do not attach adhesive labels, as this is not a sustainable solution and would require management of vast quantities of labels, given the large number of components.

Declarations of conformity

You can conveniently download and manage declarations of conformity in our MyVITA online portal. Simply register at www.vita-zahnfabrik.com/MyVITA-Registrierung

In the "Declarations of conformity" section, you can access the service "Automatic updates for your declarations of conformity." If requested, your selected VITA declarations of conformity are updated automatically. You will also be informed via email when a new version is available.